



```
Current Data Parameters
NAME                Rg1
EXPNO                1
PROCNO              1

F2 - Acquisition Parameters
Date_                20140819
Time                 13.42
INSTRUM              spect
PROBHD               5 mm PABBO BB-
PULPROG              zg30
TD                   65536
SOLVENT              DMSO
NS                   16
DS                   2
SWH                  8223.685 Hz
FIDRES               0.125483 Hz
AQ                   3.9846387 sec
RG                   109.31
DW                   60.800 usec
DE                   6.50 usec
TE                   301.0 K
D1                   1.00000000 sec
TD0                  1

===== CHANNEL f1 =====
NUC1                 1H
P1                   9.23 usec
PLW1                 20.00000000 W
SFO1                 400.1324710 MHz

F2 - Processing parameters
SI                   65536
SF                   400.1300028 MHz
WDW                  EM
SSB                  0
LB                   0.30 Hz
GB                   0
PC                   1.00
```